

Declaration of conformity to EU Medical Device Regulation 2017/745

FB V7.3.02-04-6

Level Menufestures	Traces Medical Ombili		
Legal Manufacturer	Tracoe Medical GmbH		
	Reichelsheimer Str. 1/3		
	55268 Nieder-Olm		
	Germany		
	SRN: DE-MF-000006938		
EU Product Classification according to Annex VIII	lla		
	Rule Number:		
	Dilation Set: Rule 6 / Seldinger Guide Wires: Rule 5		
Intended Purpose	The Tracoe Experc Dilation Set (REF 520) is intended for		
	performing a percutaneous dilation tracheostomy,		
	applying the Ciaglia technique with a Seldinger guide		
	wire.		
	The Tracoe Seldinger Guide Wire with Guiding Catheter		
	(REF 517) facilitates the reinsertion when changing the		
	Tracoe Tracheostomy Tubes with minimally traumatic		
	inserter.		
	The Tracoe Seldinger Guide Wire (REF 518) facilitates		
	the reinsertion when changing Tracoe Twist Plus		
	tracheostomy tubes with minimally traumatic insertion		
	system, Tracoe Vario tracheostomy tubes with minimally		
	traumatic insertion system, and all Tracoe tracheostomy		
	tubes with perforated obturator.		
Basic UDI-DI	Dilation Set: 4035324EXPERC DIL SET46		
	Seldinger Guide Wires:		
	4035324SELDI GUIDE WIREQT		
Conformity Assessment Procedure	Annex IX		
Notified Body Name and Number	TÜV Süd Product Service GmbH (0123)		
Notified Body Certificate Type and Number	EU Quality Management System Certificate		
	G10 036993 0025		
Conformity to Common Specification(s)	No relevant Common Specification to list		
Conformity to other Union Legislation(s)	No relevant Union Legislation to list		

This EU Declaration of Conformity is issued under the sole responsibility of Tracoe Medical GmbH. Tracoe Medical GmbH declares that the devices covered by the present declaration are in conformity with the Regulation (EU) 20217/745 of the European Parliament and Council of 5 April 2017 on medical devices and the requirements specified herein are fulfilled.

Date of signature:

2024-08-07

yyyy-mm-dd

Place of signature:

Nieder-Olm, Germany

Place, Country

Name, Title

Signed on behalf of Tracoe Medical GmbH:



Dr. Hans-Marcus Kränzler, Managing Director

Freigegeben: Felix Dreher

Datum: 07.05.2024 11:04

This EU Declaration of conformity is applicable for following catalogue numbers:

Catalogue Number	Product Name	UDI-DI	Original CE Marking Date yyyy-mm-dd
REF 520	Tracoe Experc Dilation Set	04035324017263	2006-12-01
REF 517	Tracoe Seldinger Guide Wire with Guiding Catheter	04035324016310	2006-12-01
REF 518	Tracoe Seldinger Guide Wire	04035324016860	2006-12-01